

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040121\
 Data File : VW020845.D
 Acq On : 01 Apr 2021 17:26
 Operator : SY/MD
 Sample : M1879-10
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW1A6

Quant Time: Apr 02 07:19:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 02 07:14:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	673950	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	640757	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	309526	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	163234	34.12	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery =	68.24%		
7) Chloroethane-d5	1.571	69	143062	39.55	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery =	79.10%		
11) 1,1-Dichloroethene-d2	2.108	63	283538	31.41	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	62.82%		
21) 2-Butanone-d5	3.899	46	265559	85.31	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery =	85.31%		
24) Chloroform-d	4.352	84	439434	46.21	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	92.42%		
26) 1,2-Dichloroethane-d4	5.037	65	301826	48.18	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	96.36%		
32) Benzene-d6	5.053	84	796615	46.64	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	93.28%		
36) 1,2-Dichloropropane-d6	6.072	67	232966	44.43	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	88.86%		
41) Toluene-d8	7.320	98	731184	45.99	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	91.98%		
43) trans-1,3-Dichloroprop...	7.625	79	125987	45.16	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	90.32%		
47) 2-Hexanone-d5	8.092	63	193581	88.06	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	88.06%		
57) 1,1,2,2-Tetrachloroeth...	10.220	84	325020	42.55	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	85.10%		
64) 1,2-Dichlorobenzene-d4	11.632	152	305688	50.89	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	101.78%		
Target Compounds						Qvalue
20) cis-1,2-Dichloroethene	3.918	96	15414	3.19	ug/L	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040121\
Data File : VV020845.D
Acq On : 01 Apr 2021 17:26
Operator : SY/MD
Sample : M1879-10
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW1A6

Quant Time: Apr 02 07:19:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 02 07:14:26 2021
Response via : Initial Calibration

