

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040121\
 Data File : VW020846.D
 Acq On : 01 Apr 2021 17:50
 Operator : SY/MD
 Sample : M1879-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW1A7

Quant Time: Apr 02 07:19:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 02 07:14:26 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	631962	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	603603	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	295221	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	167593	37.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.70%
7) Chloroethane-d5	1.571	69	144826	42.70	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	85.40%
11) 1,1-Dichloroethene-d2	2.108	63	287210	33.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.86%
21) 2-Butanone-d5	3.899	46	272879	93.49	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.49%
24) Chloroform-d	4.352	84	444918	49.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.78%
26) 1,2-Dichloroethane-d4	5.037	65	301808	51.38	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.76%
32) Benzene-d6	5.053	84	809179	50.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.58%
36) 1,2-Dichloropropane-d6	6.072	67	240948	48.78	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	97.56%
41) Toluene-d8	7.320	98	739993	49.41	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.82%
43) trans-1,3-Dichloroprop...	7.625	79	125456	47.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.48%
47) 2-Hexanone-d5	8.091	63	195998	94.64	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.64%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	331787	46.11	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.22%
64) 1,2-Dichlorobenzene-d4	11.632	152	308819	53.90	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	107.80%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040121\
Data File : VV020846.D
Acq On : 01 Apr 2021 17:50
Operator : SY/MD
Sample : M1879-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW1A7

Quant Time: Apr 02 07:19:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 02 07:14:26 2021
Response via : Initial Calibration

