

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040124\
 Data File : VV034914.D
 Acq On : 01 Apr 2024 09:40
 Operator : SY/MD
 Sample : VV0401WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK258

Quant Time: Apr 02 02:09:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 02 02:07:06 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	360472	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	343326	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.184	152	148025	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	150647	43.274	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	86.540%	
7) Chloroethane-d5	1.532	69	149843	49.556	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	99.120%	
11) 1,1-Dichloroethene-d2	2.059	65	68035	44.583	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	89.160%	
21) 2-Butanone-d5	3.793	46	188477	91.287	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	91.290%	
24) Chloroform-d	4.246	84	259802	44.387	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.780%	
26) 1,2-Dichloroethane-d4	4.940	65	168392	45.314	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.620%	
32) Benzene-d6	4.956	84	523628	45.405	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.800%	
36) 1,2-Dichloropropane-d6	5.985	67	170843	46.651	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	93.300%	
41) Toluene-d8	7.243	98	448750	43.207	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	86.420%	
43) trans-1,3-Dichloroprop...	7.554	79	71668	40.256	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	80.520%	
47) 2-Hexanone-d5	8.024	63	112703	81.398	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	81.400%	
56) 1,1,2,2-Tetrachloroeth...	10.152	84	222903	44.200	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	88.400%	
66) 1,2-Dichlorobenzene-d4	11.561	152	157949	47.525	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.060%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040124\
Data File : VV034914.D
Acq On : 01 Apr 2024 09:40
Operator : SY/MD
Sample : VV0401WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK258

Quant Time: Apr 02 02:09:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 02 02:07:06 2024
Response via : Initial Calibration

