

Data File : VV010084.D
Acq On : 03 Apr 2019 03:11
Operator : SY/MD
Sample : K2148-22
Misc : 5.88G/10mL/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Apr 03 08:21:25 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 03 06:44:57 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	259665	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	239425	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	77772	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	65706	18.90	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	75.60%
7) Chloroethane-d5	1.57	69	106203	24.60	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	98.40%
10) 1,1-Dichloroethene-d2	2.12	63	133730	15.25	ug/L	-0.01
Spiked Amount	25.000	Range	45 - 110	Recovery	=	61.00%
20) 2-Butanone-d5	3.95	46	19748m	19.96	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	39.92%
24) Chloroform-d	4.40	84	163890	25.05	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	100.20%
26) 1,2-Dichloroethane-d4	5.08	65	98257	26.85	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	107.40%
29) Benzene-d6	5.09	84	329690	26.55	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	106.20%
33) 1,2-Dichloropropane-d6	6.12	67	99755	27.22	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	108.88%
37) Toluene-d8	7.36	98	279325	23.35	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	93.40%
38) trans-1,3-Dichloropropene-	7.66	79	28455	18.17	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	72.68%
39) 2-Hexanone-d5	8.14	63	12743	17.34	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	34.68%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	78752	21.79	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	87.16%
61) 1,2-Dichlorobenzene-d4	11.67	152	78339	25.86	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	103.44%

Target Compounds					Qvalue
13) Acetone	2.20	43	2217	2.267 ug/L	71
16) Methylene chloride	2.53	84	12724	2.789 ug/L	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VV010084.D
Acq On : 03 Apr 2019 03:11
Operator : SY/MD
Sample : K2148-22
Misc : 5.88G/10mL/MSVOA_V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Apr 03 08:21:25 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 03 06:44:57 2019
Response via : Initial Calibration

