

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040220\
 Data File : VV015368.D
 Acq On : 02 Apr 2020 12:46
 Operator : SY/MD
 Sample : L2005-04 10X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CODE6

Quant Time: Apr 03 02:30:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 03 01:25:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	791546	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	752468	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	371125	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	223852	31.54	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.08%
7) Chloroethane-d5	1.58	69	203123	33.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	67.00%#
11) 1,1-Dichloroethene-d2	2.11	63	377763	29.78	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.56%#
21) 2-Butanone-d5	3.94	46	292869	77.64	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	77.64%
24) Chloroform-d	4.37	84	421121	35.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	71.88%
26) 1,2-Dichloroethane-d4	5.06	65	283993	38.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.82%
32) Benzene-d6	5.07	84	908944	39.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.44%
36) 1,2-Dichloropropane-d6	6.09	67	274737	40.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	80.42%
41) Toluene-d8	7.34	98	873600	39.42	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	78.84%#
43) trans-1,3-Dichloropropene-	7.64	79	143507	38.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.02%
47) 2-Hexanone-d5	8.12	63	241951	77.54	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.54%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	395073	40.64	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.28%
64) 1,2-Dichlorobenzene-d4	11.65	152	349239	41.37	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.74%

Target Compounds

					Ovalue
25) Chloroform	4.40	83	25533	2.393	ug/L 99
33) Benzene	5.12	78	843732	37.883	ug/L 100
51) Chlorobenzene	8.90	112	1531677	94.417	ug/L 98
62) 1,3-Dichlorobenzene	11.21	146	216181	17.303	ug/L 98
63) 1,4-Dichlorobenzene	11.30	146	1224691	96.650	ug/L 98
65) 1,2-Dichlorobenzene	11.67	146	1845562	144.825	ug/L 99
68) 1,2,4-trichlorobenzene	13.29	180	462405	61.137	ug/L 99
70) 1,2,3-Trichlorobenzene	13.77	180	188410	24.183	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

