

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040220\
 Data File : VV015380.D
 Acq On : 02 Apr 2020 17:38
 Operator : SY/MD
 Sample : L2003-04RE
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EK5RE

Quant Time: Apr 03 03:54:28 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 03 01:25:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	691832	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	660828	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	298247	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	216675	34.93	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.86%
7) Chloroethane-d5	1.58	69	196768	37.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.26%
11) 1,1-Dichloroethene-d2	2.12	63	333231	30.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	60.10%
21) 2-Butanone-d5	3.94	46	301851	91.55	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.55%
24) Chloroform-d	4.38	84	410809	40.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.22%
26) 1,2-Dichloroethane-d4	5.06	65	269642	42.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.54%
32) Benzene-d6	5.08	84	848198	41.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.36%
36) 1,2-Dichloropropane-d6	6.09	67	257819	42.97	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	85.94%
41) Toluene-d8	7.34	98	820468	42.16	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.32%
43) trans-1,3-Dichloropropene-	7.64	79	131210	39.58	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.16%
47) 2-Hexanone-d5	8.12	63	237568	86.69	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.69%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	368528	43.17	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.34%
64) 1,2-Dichlorobenzene-d4	11.65	152	299562	44.15	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.30%

Target Compounds					Ovalue
25) Chloroform	4.41	83	10985	1.178 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV040220\
Data File : VV015380.D
Acq On : 02 Apr 2020 17:38
Operator : SY/MD
Sample : L2003-04RE
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EK5RE

Quant Time: Apr 03 03:54:28 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 03 01:25:45 2020
Response via : Initial Calibration

