

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040220\
 Data File : VV015381.D
 Acq On : 02 Apr 2020 18:02
 Operator : SY/MD
 Sample : L2003-05RE
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0EK8RE

Quant Time: Apr 03 03:59:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 03 01:25:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	682281	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	662883	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	299027	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	215796	35.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	70.54%
7) Chloroethane-d5	1.58	69	200303	38.33	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	76.66%
11) 1,1-Dichloroethene-d2	2.12	63	346228	31.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	63.32%
21) 2-Butanone-d5	3.94	46	299657	92.16	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	92.16%
24) Chloroform-d	4.38	84	414815	41.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.14%
26) 1,2-Dichloroethane-d4	5.06	65	270126	42.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.88%
32) Benzene-d6	5.08	84	853997	41.83	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.66%
36) 1,2-Dichloropropane-d6	6.09	67	259007	43.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.06%
41) Toluene-d8	7.34	98	822407	42.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	84.26%
43) trans-1,3-Dichloropropene-	7.64	79	133688	40.20	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	80.40%
47) 2-Hexanone-d5	8.12	63	226229	82.30	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.30%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	371315	43.36	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	86.72%
64) 1,2-Dichlorobenzene-d4	11.65	152	302853	44.52	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.04%

Target Compounds					Ovalue
25) Chloroform	4.40	83	10857	1.180 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV040220\
Data File : VV015381.D
Acq On : 02 Apr 2020 18:02
Operator : SY/MD
Sample : L2003-05RE
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0EK8RE

Quant Time: Apr 03 03:59:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 03 01:25:45 2020
Response via : Initial Calibration

