

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040220\
 Data File : VV015391.D
 Acq On : 02 Apr 2020 23:17
 Operator : SY/MD
 Sample : L2064-12
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK02

Quant Time: Apr 03 06:23:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 03 05:25:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	648821	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	621190	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	276214	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	228091	39.21	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	78.42%
7) Chloroethane-d5	1.58	69	199526	40.15	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	80.30%
11) 1,1-Dichloroethene-d2	2.12	63	343246	33.01	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.02%
21) 2-Butanone-d5	3.94	46	297758	96.30	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	96.30%
24) Chloroform-d	4.38	84	403340	42.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.00%
26) 1,2-Dichloroethane-d4	5.06	65	264968	44.29	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.58%
32) Benzene-d6	5.08	84	846120	44.23	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.46%
36) 1,2-Dichloropropane-d6	6.10	67	255351	45.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.54%
41) Toluene-d8	7.34	98	788730	43.12	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.24%
43) trans-1,3-Dichloropropene-	7.64	79	121927	39.12	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.24%
47) 2-Hexanone-d5	8.12	63	232524	90.27	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.27%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	362695	45.19	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.38%
64) 1,2-Dichlorobenzene-d4	11.65	152	289649	46.10	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.20%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040220\
Data File : VV015391.D
Acq On : 02 Apr 2020 23:17
Operator : SY/MD
Sample : L2064-12
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK02

Quant Time: Apr 03 06:23:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 03 05:25:49 2020
Response via : Initial Calibration

