


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV040319\
Data File : VV010090.D
Acq On : 03 Apr 2019 14:07
Operator : SY/MD
Sample : K2215-04
Misc : 5.31G/10mL/MSVOA_V/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF2M5

Manual Integrations
APPROVED

MMDadoda
4/5/2019 10:40:30 AM

Quant Time: Apr 04 06:59:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 04 06:01:11 2019
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV040319\
Data File : VV010090.D
Acq On : 03 Apr 2019 14:07
Operator : SY/MD
Sample : K2215-04
Misc : 5.31G/10mL/MSVOA_V/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF2M5

Manual Integrations
APPROVED

MMDadoda
4/5/2019 10:40:30 AM

Quant Time: Apr 04 06:05:14 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 04 06:01:11 2019
Response via : Initial Calibration



(10) 1,1-Dichloroethene-d2 (S)

2.212min (+0.084) 0.03ug/L

response 239

Ion	Exp%	Act%
63.00	100	100
98.00	69.50	57.32
65.00	25.10	26.78
0.00	0.00	0.00

Quantitation Report (Qedit)


Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV040319\
 Data File : VV010090.D
 Acq On : 03 Apr 2019 14:07
 Operator : SY/MD
 Sample : K2215-04
 Misc : 5.31G/10mL/MSVOA_V/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BF2M5

Manual Integrations
 APPROVED

MMDadoda
 4/5/2019 10:40:30 AM

Quant Time: Apr 04 06:05:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 06:01:11 2019
 Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV040319\
Data File : VV010090.D
Acq On : 03 Apr 2019 14:07
Operator : SY/MD
Sample : K2215-04
Misc : 5.31G/10mL/MSVOA_V/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BF2M5

Manual Integrations
APPROVED

MMDadoda
4/5/2019 10:40:30 AM

Quant Time: Apr 04 06:05:14 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 04 06:01:11 2019
Response via : Initial Calibration



TIC: VV010090.D

(20) 2-Butanone-d5 (S)

4.068min (+0.129) 0.43ug/L

response 388

Ion	Exp%	Act%
46.00	100	100
77.00	25.70	28.61
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV040319\
 Data File : VV010090.D
 Acq On : 03 Apr 2019 14:07
 Operator : SY/MD
 Sample : K2215-04
 Misc : 5.31G/10mL/MSVOA_V/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BF2M5

Manual Integrations
 APPROVED

MMDadoda
 4/5/2019 10:40:30 AM

Quant Time: Apr 04 06:05:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 06:01:11 2019
 Response via : Initial Calibration



TIC: VV010090.D

(20) 2-Butanone-d5 (S)

3.945min (+0.006) 42.08ug/L m

response 37892

Ion	Exp%	Act%
46.00	100	100
77.00	25.70	0.29#
0.00	0.00	0.00
0.00	0.00	0.00

MD
 4/6/19

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_V\DATA\VV040319\
 Data File : VV010090.D
 Acq On : 03 Apr 2019 14:07
 Operator : SY/MD
 Sample : K2215-04
 Misc : 5.31G/10mL/MSVOA_V/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BF2M5

Manual Integrations
 APPROVED

MMDadoda
 4/5/2019 10:40:30 AM

Quant Time: Apr 04 06:59:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 04 06:01:11 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev (Min)
1) 1,4-Difluorobenzene	5.66	114	236385	25.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	225547	25.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	101742	25.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	57352	18.12	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	72.48%
7) Chloroethane-d5	1.57	69	72818	18.53	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	74.12%
10) 1,1-Dichloroethene-d2	2.13	63	120474m	15.09	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	60.36%
20) 2-Butanone-d5	3.95	46	37892m	42.08	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.16%
24) Chloroform-d	4.40	84	129323	21.72	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	86.88%
26) 1,2-Dichloroethane-d4	5.08	65	76176	22.86	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	91.44%
29) Benzene-d6	5.10	84	266676	22.80	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	91.20%
33) 1,2-Dichloropropane-d6	6.12	67	79566	23.05	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	92.20%
37) Toluene-d8	7.36	98	240347	21.33	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	85.32%
38) trans-1,3-Dichloropropene-	7.66	79	27984	18.96	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	75.84%
39) 2-Hexanone-d5	8.14	63	27583	39.84	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.68%
48) 1,1,2,2-Tetrachloroethane-	10.26	84	71805	21.09	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	84.36%
61) 1,2-Dichlorobenzene-d4	11.67	152	93698	23.64	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	94.56%

MD
 4/6/19

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
13) Acetone	2.20	43	2786	3.129	ug/L	90
16) Methylene chloride	2.54	84	5053	1.217	ug/L	97
47) Tetrachloroethene	8.02	164	107136	40.317	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed