

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040320\
 Data File : VV015404.D
 Acq On : 03 Apr 2020 18:31
 Operator : SY/MD
 Sample : L2065-21
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Apr 04 02:32:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 04 00:09:13 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	788494	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	744549	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	351448	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	262026	37.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	74.12%
7) Chloroethane-d5	1.58	69	234023	38.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.50%
11) 1,1-Dichloroethene-d2	2.12	63	376169	29.77	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.54%#
21) 2-Butanone-d5	3.94	46	314165	83.60	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	83.60%
24) Chloroform-d	4.38	84	456112	39.08	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	78.16%
26) 1,2-Dichloroethane-d4	5.06	65	294220	40.47	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.94%
32) Benzene-d6	5.08	84	936559	40.85	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.70%
36) 1,2-Dichloropropane-d6	6.10	67	282120	41.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.46%
41) Toluene-d8	7.34	98	906520	41.34	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.68%
43) trans-1,3-Dichloropropene-	7.64	79	145383	38.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.84%
47) 2-Hexanone-d5	8.12	63	237273	76.85	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.85%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	397167	41.29	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.58%
64) 1,2-Dichlorobenzene-d4	11.65	152	335669	41.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.98%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV040320\
Data File : VV015404.D
Acq On : 03 Apr 2020 18:31
Operator : SY/MD
Sample : L2065-21
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Apr 04 02:32:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM040120WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 04 00:09:13 2020
Response via : Initial Calibration

