


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV040519\
Data File : VV010123.D
Acq On : 05 Apr 2019 08:48
Operator : SY/MD
Sample : BFB85
Misc : 5.00G/10mL/MSVOA V/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB85

Integration File: LSCINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOM2VLM040219S.M
Title : VOC Analysis
Last Update : Sun Apr 07 05:35:34 2019



Spectrum Information: Average of 10.116 to 10.122 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.8	22525	PASS
75	95	30	80	52.7	62984	PASS
95	95	100	100	100.0	119549	PASS
96	95	5	9	7.0	8335	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	88.2	105432	PASS
175	174	5	9	7.9	8366	PASS
176	174	95	101	95.5	100696	PASS
177	176	5	9	7.0	7084	PASS

Scan 2795 (10.077 min): VV010123.D
BFB85

m/z	Abundance
39.90	109.0
44.00	248.0
46.30	105.0
117.20	147.0

Instrument :
MSVOA_V
ClientSampleId :
BFB85

m/z	Abundance
36.00	1285.0
37.00	5855.0
38.00	5220.0
39.00	2348.0
39.90	203.0
44.10	1160.0
45.00	928.0
46.10	182.0
46.90	1029.0
48.00	697.0
49.00	5362.0
50.00	21848.0
51.00	6393.0
52.00	224.0
54.80	357.0
56.00	2575.0
57.00	3960.0
57.90	222.0
60.00	1709.0
61.00	6105.0
62.00	5327.0
63.00	5267.0
64.10	586.0
65.00	754.0
66.20	150.0
66.90	306.0
68.00	11591.0
69.00	12570.0
70.00	1097.0
72.00	670.0
73.00	6212.0
74.00	21968.0
75.00	60152.0
76.00	5633.0
77.10	663.0
77.90	450.0
79.00	5928.0
79.90	1698.0
80.90	6588.0
81.90	1478.0
82.90	125.0
85.90	193.0
87.00	3975.0
88.00	4788.0
90.90	580.0
92.00	3660.0
93.00	5692.0
94.00	14927.0
95.00	114968.0
96.00	8151.0
97.00	323.0
102.90	111.0
104.00	691.0
105.00	269.0
105.80	861.0
106.80	155.0
110.00	123.0
110.80	353.0
112.00	122.0
112.90	221.0
114.90	209.0
115.80	622.0
116.90	906.0
117.80	521.0
118.90	897.0
124.00	109.0
127.90	489.0
129.10	230.0
129.90	483.0
130.80	139.0
131.20	108.0
134.90	200.0
136.90	259.0
140.90	1816.0
141.90	262.0
142.90	1811.0
144.80	203.0
145.90	224.0
147.80	344.0
149.90	136.0
151.80	169.0
153.60	101.0
154.90	325.0
156.80	280.0
158.90	198.0
160.90	236.0
168.90	210.0
169.60	108.0
171.20	280.0
171.90	1022.0
174.00	99088.0
175.00	8379.0

Instrument :
MSVOA_V
ClientSampleId :
BFB85

176.00	94216.0
177.00	6892.0
178.10	218.0

Instrument :
MSVOA_V
ClientSampleId :
BFB85

m/z	Abundance
36.00	1323.0
37.00	6094.0
38.00	5628.0
39.00	2406.0
39.90	228.0
44.10	1246.0
45.00	1140.0
46.10	175.0
46.90	1046.0
48.00	775.0
49.00	5451.0
50.00	23424.0
51.00	6647.0
52.10	271.0
54.80	408.0
56.00	2559.0
57.00	4094.0
57.90	190.0
60.00	1645.0
61.00	6703.0
62.00	6008.0
63.00	5519.0
64.00	613.0
65.00	900.0
66.20	150.0
67.00	300.0
68.00	12589.0
69.00	13743.0
70.00	1162.0
71.90	716.0
73.00	6330.0
74.00	23912.0
75.00	65024.0
76.00	6309.0
77.00	702.0
78.00	603.0
78.90	6291.0
79.90	1867.0
80.90	6846.0
81.90	1524.0
82.90	124.0
86.00	243.0
87.00	4129.0
87.90	4889.0
91.00	579.0
92.10	4095.0
93.00	6103.0
94.00	15347.0
95.00	123488.0
96.00	8472.0
97.00	284.0
102.90	130.0
104.00	766.0
105.00	274.0
105.80	950.0
107.00	168.0
110.00	102.0
110.80	328.0
112.00	114.0
112.90	234.0
114.80	266.0
115.80	628.0
116.90	978.0
117.80	617.0
118.90	880.0
125.90	106.0
127.80	555.0
129.10	286.0
129.90	490.0
130.70	127.0
135.10	248.0
137.00	335.0
140.90	1886.0
141.90	289.0
142.90	1942.0
144.00	124.0
144.80	282.0
145.80	286.0
147.00	112.0
147.80	353.0
149.90	112.0
151.80	153.0
154.90	363.0
156.90	282.0
158.90	189.0
160.90	236.0
168.90	238.0
169.60	106.0
169.90	102.0
171.20	322.0
172.00	1082.0
174.00	108392.0

Instrument :
MSVOA_V
ClientSampleId :
BFB85

175.00	8583.0
176.00	103992.0
177.00	7391.0
177.90	148.0

Instrument :
MSVOA_V
ClientSampleId :
BFB85

m/z	Abundance
36.00	1130.0
37.00	5636.0
38.00	5411.0
39.00	2172.0
39.90	219.0
40.90	108.0
44.00	1158.0
45.00	1146.0
46.00	146.0
46.90	1086.0
47.90	857.0
49.00	4907.0
50.00	22304.0
51.00	6525.0
52.10	303.0
54.90	381.0
56.00	2394.0
57.00	4004.0
58.10	177.0
60.10	1349.0
61.00	6728.0
62.00	5977.0
63.00	5134.0
64.00	599.0
65.10	936.0
66.20	120.0
66.80	263.0
67.10	266.0
68.00	12542.0
69.00	13545.0
70.00	1038.0
72.00	713.0
73.00	5631.0
74.00	23144.0
75.00	63776.0
76.10	6005.0
76.90	786.0
78.00	642.0
78.90	6404.0
79.90	1915.0
80.90	6414.0
82.00	1467.0
82.90	107.0
86.00	262.0
87.00	4156.0
88.00	4534.0
91.00	575.0
92.00	4081.0
93.00	5921.0
94.00	14554.0
95.00	120192.0
96.00	8383.0
97.20	234.0
104.00	750.0
105.00	240.0
105.90	915.0
107.00	184.0
110.80	289.0
112.00	109.0
113.00	212.0
114.80	273.0
115.80	601.0
116.90	933.0
117.80	675.0
118.80	801.0
125.80	130.0
127.80	599.0
129.10	250.0
130.00	447.0
130.80	148.0
135.00	249.0
137.00	354.0
138.70	119.0
140.90	1885.0
141.80	270.0
142.90	2024.0
144.00	114.0
144.90	315.0
145.90	327.0
147.00	139.0
147.90	381.0
151.80	125.0
154.90	393.0
157.10	250.0
158.90	172.0
160.80	181.0
168.90	182.0
169.50	106.0
172.00	1059.0
174.00	108816.0
175.00	8136.0
176.00	103880.0

Instrument :
MSVOA_V
ClientSampleId :
BFB85

177.00	6970.0
177.90	160.0

Instrument :
MSVOA_V
ClientSampleId :
BFB85