

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040521\
 Data File : VV020909.D
 Acq On : 05 Apr 2021 12:56
 Operator : SY/MD
 Sample : VV0405MBL01
 Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK89

Quant Time: Apr 06 06:22:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 06 06:22:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	636742	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	614117	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	314874	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	181969	40.25	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	80.50%
7) Chloroethane-d5	1.568	69	149048	43.61	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.22%
11) 1,1-Dichloroethene-d2	2.108	63	278239	32.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.24%
21) 2-Butanone-d5	3.899	46	238018	80.93	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	80.93%
24) Chloroform-d	4.349	84	426818	47.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.00%
26) 1,2-Dichloroethane-d4	5.034	65	296707	50.13	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.26%
32) Benzene-d6	5.050	84	798033	48.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.50%
36) 1,2-Dichloropropane-d6	6.069	67	233863	46.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.06%
41) Toluene-d8	7.317	98	764740	50.19	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.38%
43) trans-1,3-Dichloroprop...	7.622	79	132138	49.42	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	98.84%
47) 2-Hexanone-d5	8.092	63	172629	81.93	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.93%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	308444	42.13	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.26%
64) 1,2-Dichlorobenzene-d4	11.628	152	324436	53.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.18%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040521\
Data File : VV020909.D
Acq On : 05 Apr 2021 12:56
Operator : SY/MD
Sample : VV0405MBL01
Misc : 5.0g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK89

Quant Time: Apr 06 06:22:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 06 06:22:01 2021
Response via : Initial Calibration

