

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040521\
 Data File : VW020931.D
 Acq On : 05 Apr 2021 21:43
 Operator : SY/MD
 Sample : M1880-09
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW1D6

Quant Time: Apr 06 06:28:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 06 06:22:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	595840	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	563519	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	285033	50.00	ug/L	0.00

System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	145634	34.43	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	68.86%
7) Chloroethane-d5	1.571	69	123006	38.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	76.92%
11) 1,1-Dichloroethene-d2	2.108	63	237478	29.76	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.52%#
21) 2-Butanone-d5	3.902	46	188794	68.60	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	68.60%
24) Chloroform-d	4.352	84	380524	45.26	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.52%
26) 1,2-Dichloroethane-d4	5.034	65	259984	46.94	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.88%
32) Benzene-d6	5.053	84	677395	45.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.20%
36) 1,2-Dichloropropane-d6	6.072	67	199903	43.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.70%
41) Toluene-d8	7.320	98	649616	46.46	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.92%
43) trans-1,3-Dichloroprop...	7.625	79	105177	42.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.74%
47) 2-Hexanone-d5	8.091	63	141604	73.24	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.24%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	275913	41.07	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.14%
64) 1,2-Dichlorobenzene-d4	11.631	152	280434	50.69	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.38%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040521\
Data File : VV020931.D
Acq On : 05 Apr 2021 21:43
Operator : SY/MD
Sample : M1880-09
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW1D6

Quant Time: Apr 06 06:28:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 06 06:22:01 2021
Response via : Initial Calibration

