

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040523\
 Data File : VV030404.D
 Acq On : 05 Apr 2023 16:34
 Operator : SY/MD
 Sample : 02194-03DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHDW0DL

Quant Time: Apr 06 05:27:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Mar 31 05:37:21 2023
 Response via : Initial Calibration

Compound		R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards							
1) 1,4-Difluorobenzene		5.545	114	130804	5.000	ug/L	0.00
28) Chlorobenzene-d5		8.792	117	126393	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4		11.194	152	63768	5.000	ug/L	0.00
System Monitoring Compounds							
4) Vinyl Chloride-d3		1.282	65	43651	5.039	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	100.800%	
7) Chloroethane-d5		1.536	69	43689	5.197	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	104.000%	
11) 1,1-Dichloroethene-d2		2.063	63	90073	4.382	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.600%	
20) 2-Butanone-d5		3.828	46	110118	49.178	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.360%	
24) Chloroform-d		4.262	84	105266	5.020	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.400%	
26) 1,2-Dichloroethane-d4		4.953	65	53248	5.383	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.600%	
32) Benzene-d6		4.969	84	191642	5.381	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	107.600%	
36) 1,2-Dichloropropane-d6		5.998	67	59530	5.339	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.800%	
41) Toluene-d8		7.252	98	161084	4.925	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.400%	
43) trans-1,3-Dichloroprop...		7.564	79	19623	4.674	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.400%	
46) 2-Hexanone-d5		8.037	63	84924	50.340	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	100.680%	
56) 1,1,2,2-Tetrachloroeth...		10.162	84	42692	4.858	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	97.200%	
66) 1,2-Dichlorobenzene-d4		11.570	152	62021	5.442	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.800%	
Target Compounds							Qvalue
6) Bromomethane		1.497	94	618	0.142	ug/L	80
12) 1,1-Dichloroethene		2.076	96	25044	2.291	ug/L #	77
19) 1,1-Dichloroethane		3.121	63	21831	0.978	ug/L	94
22) cis-1,2-Dichloroethene		3.825	96	55386	4.793	ug/L #	99
29) 1,1,1-Trichloroethane		4.529	97	17304	0.852	ug/L	97
34) Trichloroethene		5.844	95	34039	2.811	ug/L	99
47) Tetrachloroethene		7.915	164	97050	9.082	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040523\
Data File : VV030404.D
Acq On : 05 Apr 2023 16:34
Operator : SY/MD
Sample : 02194-03DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHDW0DL

Quant Time: Apr 06 05:27:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Mar 31 05:37:21 2023
Response via : Initial Calibration

