

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040524\
 Data File : VV034994.D
 Acq On : 05 Apr 2024 10:36
 Operator : SY/MD
 Sample : P1987-11
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MCORG9

Quant Time: Apr 06 01:01:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 06 00:54:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	349870	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	344997	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	145667	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	125988	37.287	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	74.580%	
7) Chloroethane-d5	1.532	69	127732	43.523	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	87.040%	
11) 1,1-Dichloroethene-d2	2.060	65	60588	40.906	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	81.820%	
21) 2-Butanone-d5	3.796	46	166434	83.053	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	83.050%	
24) Chloroform-d	4.246	84	257053	45.248	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.500%	
26) 1,2-Dichloroethane-d4	4.941	65	165662	45.930	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	91.860%	
32) Benzene-d6	4.957	84	507185	43.766	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	87.540%	
36) 1,2-Dichloropropane-d6	5.986	67	169932	46.178	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	92.360%	
41) Toluene-d8	7.243	98	428835	41.090	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	82.180%	
43) trans-1,3-Dichloroprop...	7.551	79	68465	38.270	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	76.540%	
47) 2-Hexanone-d5	8.027	63	87556	62.930	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	62.930%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	218186	43.055	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	86.120%	
66) 1,2-Dichlorobenzene-d4	11.558	152	156653	47.898	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	95.800%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040524\
Data File : VV034994.D
Acq On : 05 Apr 2024 10:36
Operator : SY/MD
Sample : P1987-11
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MC0RG9

Quant Time: Apr 06 01:01:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 06 00:54:20 2024
Response via : Initial Calibration

