

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040524\
 Data File : VW035008.D
 Acq On : 05 Apr 2024 16:07
 Operator : SY/MD
 Sample : P2007-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH2P6

Quant Time: Apr 06 01:04:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 06 00:54:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	379369	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	369228	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	161056	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	108851	29.710	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	59.420%#	
7) Chloroethane-d5	1.529	69	116379	36.572	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	73.140%	
11) 1,1-Dichloroethene-d2	2.060	65	52640	32.777	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	65.560%	
21) 2-Butanone-d5	3.789	46	176288	81.130	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	81.130%	
24) Chloroform-d	4.249	84	243824	39.582	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	79.160%	
26) 1,2-Dichloroethane-d4	4.941	65	157170	40.187	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	80.380%	
32) Benzene-d6	4.960	84	458150	36.940	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	73.880%	
36) 1,2-Dichloropropane-d6	5.985	67	161489	41.004	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	82.000%	
41) Toluene-d8	7.243	98	381523	34.157	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	68.320%#	
43) trans-1,3-Dichloroprop...	7.555	79	64070	33.463	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	66.920%	
47) 2-Hexanone-d5	8.024	63	99703	66.958	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	66.960%	
56) 1,1,2,2-Tetrachloroeth...	10.149	84	221334	40.810	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	81.620%	
66) 1,2-Dichlorobenzene-d4	11.558	152	151101	41.786	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	83.580%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040524\
Data File : VV035008.D
Acq On : 05 Apr 2024 16:07
Operator : SY/MD
Sample : P2007-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH2P6

Quant Time: Apr 06 01:04:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 06 00:54:20 2024
Response via : Initial Calibration

