

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040524\
 Data File : VW035011.D
 Acq On : 05 Apr 2024 17:19
 Operator : SY/MD
 Sample : P2006-01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MC0R44

Quant Time: Apr 06 01:05:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 06 00:54:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	359545	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	348300	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	149303	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	173959	50.099	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	= 100.200%		
7) Chloroethane-d5	1.529	69	163911	54.348	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	= 108.700%		
11) 1,1-Dichloroethene-d2	2.060	65	74131	48.703	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 97.400%		
21) 2-Butanone-d5	3.789	46	195164	94.769	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	= 94.770%		
24) Chloroform-d	4.246	84	276240	47.317	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 94.640%		
26) 1,2-Dichloroethane-d4	4.940	65	178065	48.040	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 96.080%		
32) Benzene-d6	4.960	84	571759	48.870	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	= 97.740%		
36) 1,2-Dichloropropane-d6	5.985	67	183790	49.470	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	= 98.940%		
41) Toluene-d8	7.243	98	485414	46.070	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 92.140%		
43) trans-1,3-Dichloroprop...	7.555	79	71502	39.589	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	= 79.180%		
47) 2-Hexanone-d5	8.024	63	103906	73.973	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	= 73.970%		
56) 1,1,2,2-Tetrachloroeth...	10.149	84	238567	46.631	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	= 93.260%		
66) 1,2-Dichlorobenzene-d4	11.561	152	167966	50.106	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	= 100.220%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040524\
Data File : VV035011.D
Acq On : 05 Apr 2024 17:19
Operator : SY/MD
Sample : P2006-01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MC0R44

Quant Time: Apr 06 01:05:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 06 00:54:20 2024
Response via : Initial Calibration

