

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040621\
 Data File : VW020952.D
 Acq On : 06 Apr 2021 15:47
 Operator : SY/MD
 Sample : M1880-16
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 EW1E4

Quant Time: Apr 07 02:14:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 07 02:10:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	532774	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	518318	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	258951	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	143068	37.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.64%
7) Chloroethane-d5	1.571	69	120510	42.14	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.28%
11) 1,1-Dichloroethene-d2	2.108	63	238917	33.48	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	66.96%
21) 2-Butanone-d5	3.905	46	201128	81.73	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.73%
24) Chloroform-d	4.352	84	387066	51.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.96%
26) 1,2-Dichloroethane-d4	5.037	65	270595	54.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	109.28%
32) Benzene-d6	5.053	84	701115	50.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	101.50%
36) 1,2-Dichloropropane-d6	6.072	67	201521	47.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.02%
41) Toluene-d8	7.320	98	675940	52.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.12%
43) trans-1,3-Dichloroprop...	7.625	79	112040	49.65	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	99.30%
47) 2-Hexanone-d5	8.095	63	149651	84.15	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.15%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	285498	46.21	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	92.42%
64) 1,2-Dichlorobenzene-d4	11.632	152	294689	58.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	117.28%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040621\
Data File : VV020952.D
Acq On : 06 Apr 2021 15:47
Operator : SY/MD
Sample : M1880-16
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
EW1E4

Quant Time: Apr 07 02:14:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 07 02:10:23 2021
Response via : Initial Calibration

