

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040621\
 Data File : VW020955.D
 Acq On : 06 Apr 2021 16:59
 Operator : SY/MD
 Sample : M1845-04
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Apr 07 02:15:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 07 02:10:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	612290	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	595019	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	299771	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	145922	33.57	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	67.14%	
7) Chloroethane-d5	1.571	69	122714	37.34	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	74.68%	
11) 1,1-Dichloroethene-d2	2.108	63	234688	28.62	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	57.24%#	
21) 2-Butanone-d5	3.905	46	207440	73.35	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	73.35%	
24) Chloroform-d	4.352	84	383300	44.36	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	88.72%	
26) 1,2-Dichloroethane-d4	5.037	65	269703	47.39	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	94.78%	
32) Benzene-d6	5.053	84	687027	43.32	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	86.64%	
36) 1,2-Dichloropropane-d6	6.072	67	199284	40.93	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	81.86%	
41) Toluene-d8	7.320	98	657036	44.51	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	89.02%	
43) trans-1,3-Dichloroprop...	7.625	79	110264	42.56	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	85.12%	
47) 2-Hexanone-d5	8.095	63	149267	73.12	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	73.12%	
57) 1,1,2,2-Tetrachloroeth...	10.220	84	278832	39.31	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	78.62%	
64) 1,2-Dichlorobenzene-d4	11.632	152	292607	50.29	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	100.58%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040621\
Data File : VV020955.D
Acq On : 06 Apr 2021 16:59
Operator : SY/MD
Sample : M1845-04
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Apr 07 02:15:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 07 02:10:23 2021
Response via : Initial Calibration

