

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040821\
 Data File : VV020963.D
 Acq On : 08 Apr 2021 11:26
 Operator : SY/MD
 Sample : VV0408WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK93

Quant Time: Apr 09 06:30:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Apr 09 06:30:02 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	557817	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	537107	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.255	152	262359	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	174492	44.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	88.12%
7) Chloroethane-d5	1.568	69	135696	45.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.64%
11) 1,1-Dichloroethene-d2	2.108	63	269001	36.00	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.00%
21) 2-Butanone-d5	3.892	46	182511	70.84	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	70.84%
24) Chloroform-d	4.349	84	381305	48.44	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.88%
26) 1,2-Dichloroethane-d4	5.034	65	260194	50.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.36%
32) Benzene-d6	5.050	84	694788	48.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.06%
36) 1,2-Dichloropropane-d6	6.072	67	198554	45.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.34%
41) Toluene-d8	7.317	98	674526	50.62	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.24%
43) trans-1,3-Dichloroprop...	7.625	79	112004	47.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.78%
47) 2-Hexanone-d5	8.091	63	133710	72.56	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.56%
57) 1,1,2,2-Tetrachloroeth...	10.220	84	249408	38.95	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.90%
64) 1,2-Dichlorobenzene-d4	11.631	152	286112	56.19	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	112.38%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040821\
Data File : VV020963.D
Acq On : 08 Apr 2021 11:26
Operator : SY/MD
Sample : VV0408WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK93

Quant Time: Apr 09 06:30:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SOMVLM032321WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Apr 09 06:30:02 2021
Response via : Initial Calibration

