

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040824\
 Data File : VW035048.D
 Acq On : 08 Apr 2024 12:40
 Operator : SY/MD
 Sample : P2017-06 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH5E9

Quant Time: Apr 08 13:04:54 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 06 02:25:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	403867	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	384357	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	171583	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	173727	44.541	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	89.080%
7) Chloroethane-d5	1.526	69	85031	25.100	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	50.200%#
11) 1,1-Dichloroethene-d2	2.053	65	77643	45.413	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.820%
21) 2-Butanone-d5	3.796	46	226176	97.775	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.780%
24) Chloroform-d	4.246	84	306013	46.664	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.320%
26) 1,2-Dichloroethane-d4	4.937	65	197675	47.478	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.960%
32) Benzene-d6	4.957	84	631699	48.928	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.860%
36) 1,2-Dichloropropane-d6	5.986	67	204217	49.812	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	99.620%
41) Toluene-d8	7.240	98	544267	46.810	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.620%
43) trans-1,3-Dichloroprop...	7.551	79	90861	45.588	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.180%
47) 2-Hexanone-d5	8.034	63	140655	90.742	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.740%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	269719	47.774	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.540%
66) 1,2-Dichlorobenzene-d4	11.561	152	191378	49.677	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.360%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040824\
Data File : VV035048.D
Acq On : 08 Apr 2024 12:40
Operator : SY/MD
Sample : P2017-06 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH5E9

Quant Time: Apr 08 13:04:54 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 06 02:25:24 2024
Response via : Initial Calibration

