

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040824\
 Data File : VV035042.D
 Acq On : 08 Apr 2024 10:15
 Operator : SY/MD
 Sample : VV0408WBL01
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK263

Quant Time: Apr 08 12:08:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 06 02:25:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	408634	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	385676	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.181	152	173404	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	188166	47.681	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	95.360%
7) Chloroethane-d5	1.532	69	163150	47.597	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.200%
11) 1,1-Dichloroethene-d2	2.060	65	81878	47.331	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.660%
21) 2-Butanone-d5	3.789	46	230242	98.372	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.370%
24) Chloroform-d	4.243	84	309938	46.711	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.420%
26) 1,2-Dichloroethane-d4	4.937	65	202799	48.141	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.280%
32) Benzene-d6	4.957	84	641220	49.496	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.000%
36) 1,2-Dichloropropane-d6	5.986	67	210683	51.213	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.420%
41) Toluene-d8	7.240	98	566134	48.524	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.040%
43) trans-1,3-Dichloroprop...	7.551	79	90878	45.440	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.880%
47) 2-Hexanone-d5	8.024	63	129056	82.974	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.970%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	272483	48.099	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	96.200%
66) 1,2-Dichlorobenzene-d4	11.558	152	193524	49.707	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040824\
Data File : VV035042.D
Acq On : 08 Apr 2024 10:15
Operator : SY/MD
Sample : VV0408WBL01
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK263

Quant Time: Apr 08 12:08:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 06 02:25:24 2024
Response via : Initial Calibration

