

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040824\
 Data File : VW035060.D
 Acq On : 08 Apr 2024 17:28
 Operator : SY/MD
 Sample : P2006-12
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MCORF1

Quant Time: Apr 09 02:16:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Apr 06 02:25:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	360947	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	346230	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	147072	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	146526	42.034	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.060%
7) Chloroethane-d5	1.532	69	141622	46.775	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.560%
11) 1,1-Dichloroethene-d2	2.056	65	66935	43.805	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.600%
21) 2-Butanone-d5	3.793	46	202790	98.090	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	98.090%
24) Chloroform-d	4.246	84	273397	46.648	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.300%
26) 1,2-Dichloroethane-d4	4.941	65	178967	48.096	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	96.200%
32) Benzene-d6	4.957	84	553117	47.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.120%
36) 1,2-Dichloropropane-d6	5.986	67	184805	50.041	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.080%
41) Toluene-d8	7.243	98	469826	44.857	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.720%
43) trans-1,3-Dichloroprop...	7.551	79	76097	42.385	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.760%
47) 2-Hexanone-d5	8.024	63	115617	82.802	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.800%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	241140	47.415	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.840%
66) 1,2-Dichlorobenzene-d4	11.558	152	167602	50.756	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.520%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040824\
Data File : VV035060.D
Acq On : 08 Apr 2024 17:28
Operator : SY/MD
Sample : P2006-12
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MCORF1

Quant Time: Apr 09 02:16:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Apr 06 02:25:24 2024
Response via : Initial Calibration

