

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
 Data File : VV020978.D
 Acq On : 08 Apr 2021 18:10
 Operator : SY/MD
 Sample : VSTD0.533
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.5305

Quant Time: Apr 09 07:59:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 09 07:58:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	6236	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	6011	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	2642	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	2971	0.41	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.042	65	2145	0.52	ug/L	0.00
7) 1,2-Dichloropropane-d6	6.069	67	2888	0.58	ug/L	0.00
8) Toluene-d8	7.317	98	5578	0.37	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	1762	0.66	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	5065	0.71	ug/L	97
6) Trichloroethene	5.925	95	2889	0.58	ug/L	93
9) 1,2-Dibromoethane	8.369	107	1673	0.59	ug/L #	98
12) 1,2,3-Trichloropropane	10.267	75	1495	0.67	ug/L	99
13) 1,2-Dibromo-3-chloropr...	12.427	75	198	0.65	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
Data File : VV020978.D
Acq On : 08 Apr 2021 18:10
Operator : SY/MD
Sample : VSTD0.533
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.5305

Quant Time: Apr 09 07:59:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 09 07:58:08 2021
Response via : Initial Calibration

