

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
 Data File : VV020985.D
 Acq On : 09 Apr 2021 11:15
 Operator : SY/MD
 Sample : M1906-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 X3307

Quant Time: Apr 10 00:50:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 10 00:49:57 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	5423	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	5088	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	2282	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	2718	0.53	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	106.00%
4) 1,2-Dichloroethane-d4	5.042	65	1917	0.53	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	106.00%
7) 1,2-Dichloropropane-d6	6.068	67	2563	0.55	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	110.00%
8) Toluene-d8	7.317	98	4603	0.50	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.00%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	1514	0.51	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	102.00%
Target Compounds						
						Qvalue
3) Vinyl chloride	1.303	62	15024	1.73	ug/L	92
6) Trichloroethene	5.925	95	1108	0.23	ug/L	91
9) 1,2-Dibromoethane	8.369	107	753	0.27	ug/L #	94
12) 1,2,3-Trichloropropane	10.267	75	4640	1.88	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.427	75	337	0.96	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
Data File : VV020985.D
Acq On : 09 Apr 2021 11:15
Operator : SY/MD
Sample : M1906-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
X3307

Quant Time: Apr 10 00:50:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 10 00:49:57 2021
Response via : Initial Calibration

