

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\
 Method File : SFAMVSIM040921.M
 Title : TRACE VOA SOM01.0
 Last Update : Sat Apr 10 00:49:57 2021
 Response Via : Initial Calibration

Calibration Files

0.5 =VV020983.D 0.05=VV020976.D 0.1 =VV020977.D 1.0 =VV020979.D 2 =VV020980.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD
1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.476	0.534	0.454	0.455	0.426	0.469	8.59
3) T	Vinyl chloride	0.812	0.829	0.858	0.774	0.727	0.800	6.36
4) S	1,2-Dichloroetha...	0.344	0.349	0.318	0.338	0.323	0.335	3.98
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.481	0.499	0.509	0.460	0.440	0.478	5.85
7) S	1,2-Dichloroprop...	0.480	0.466	0.418	0.478	0.438	0.456	5.97
8) S	Toluene-d8	0.928	0.974	0.860	0.915	0.891	0.914	4.66
9) T	1,2-Dibromoethane	0.278	0.258	0.279	0.276	0.262	0.271	3.60
10) S	1,1,2,2-Tetrachl...	0.293	0.320	0.279	0.289	0.276	0.291	6.07
11) I	1,4-Dichlorobenzen...	-----ISTD-----						
12)	1,2,3-Trichlorop...	0.566	0.495	0.567	0.553	0.521	0.541	5.81
13) T	1,2-Dibromo-3-ch...	0.075	0.104	0.069	0.069	0.067	0.077	20.19

(#) = Out of Range