

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
 Data File : VV020976.D
 Acq On : 08 Apr 2021 17:24
 Operator : SY/MD
 Sample : VSTD0.0531
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.05303

Quant Time: Apr 09 07:58:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 09 07:58:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	6354	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	6118	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	2685	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	339	0.05	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.066	65	222	0.05	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.092	67	285	0.06	ug/L	0.02
8) Toluene-d8	7.332	98	596	0.04	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.216	84	196	0.07	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.303	62	527	0.07	ug/L #	53
6) Trichloroethene	5.925	95	305	0.06	ug/L	84
9) 1,2-Dibromoethane	8.368	107	158	0.06	ug/L #	88
12) 1,2,3-Trichloropropane	10.284	75	133	0.06	ug/L #	88
13) 1,2-Dibromo-3-chloropr...	12.426	75	28	0.09	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
Data File : VV020976.D
Acq On : 08 Apr 2021 17:24
Operator : SY/MD
Sample : VSTD0.0531
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.05303

Quant Time: Apr 09 07:58:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 09 07:58:08 2021
Response via : Initial Calibration

