

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
 Data File : VV020980.D
 Acq On : 08 Apr 2021 18:57
 Operator : SY/MD
 Sample : VSTD00235
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002301

Quant Time: Apr 09 07:59:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 09 07:58:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Difluorobenzene	5.639	114	6379	0.50 ug/L	0.00
5) Chlorobenzene-d5	8.864	117	6203	0.50 ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	2734	0.50 ug/L	0.00
System Monitoring Compounds					
2) Vinyl Chloride-d3	1.303	65	10863	1.47 ug/L	0.00
4) 1,2-Dichloroethane-d4	5.043	65	8252	1.94 ug/L	0.00
7) 1,2-Dichloropropane-d6	6.069	67	10873	2.12 ug/L	0.00
8) Toluene-d8	7.317	98	22102	1.44 ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.216	84	6842	2.50 ug/L	0.00
Target Compounds					
3) Vinyl chloride	1.310	62	18554	2.53 ug/L	Qvalue 99
6) Trichloroethene	5.926	95	10925	2.13 ug/L	94
9) 1,2-Dibromoethane	8.353	107	6501	2.24 ug/L #	95
12) 1,2,3-Trichloropropane	10.267	75	5700	2.48 ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.427	75	733	2.33 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
Data File : VV020980.D
Acq On : 08 Apr 2021 18:57
Operator : SY/MD
Sample : VSTD00235
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002301

Quant Time: Apr 09 07:59:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 09 07:58:08 2021
Response via : Initial Calibration

