

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
 Data File : VV020981.D
 Acq On : 08 Apr 2021 19:43
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV301

Quant Time: Apr 09 08:17:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 09 08:14:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	5572	0.50	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	5349	0.50	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	2359	0.50	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	2850	0.55	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	110.00%
4) 1,2-Dichloroethane-d4	5.042	65	1996	0.54	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	108.00%
7) 1,2-Dichloropropane-d6	6.068	67	2716	0.56	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	112.00%
8) Toluene-d8	7.317	98	5051	0.52	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	104.00%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	1612	0.52	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	104.00%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.303	62	4643	0.52	ug/L	96
6) Trichloroethene	5.925	95	2626	0.51	ug/L	93
9) 1,2-Dibromoethane	8.369	107	1527	0.53	ug/L #	98
12) 1,2,3-Trichloropropane	10.267	75	1340	0.53	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.427	75	184	0.51	ug/L	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040921\
Data File : VV020981.D
Acq On : 08 Apr 2021 19:43
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV301

Quant Time: Apr 09 08:17:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM040921.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Apr 09 08:14:47 2021
Response via : Initial Calibration

