

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040924\
 Data File : VV035073.D
 Acq On : 09 Apr 2024 11:27
 Operator : SY/MD
 Sample : P2017-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BH5E9

Quant Time: Apr 10 04:55:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 10 04:53:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	307963	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	296155	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	128389	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	125285	42.124	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.240%
7) Chloroethane-d5	1.523	69	112944	43.722	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	87.440%
11) 1,1-Dichloroethene-d2	2.056	65	57145	43.832	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.660%
21) 2-Butanone-d5	3.783	46	187258	106.160	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.160%
24) Chloroform-d	4.246	84	249616	49.918	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.840%
26) 1,2-Dichloroethane-d4	4.937	65	163707	51.564	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.120%
32) Benzene-d6	4.957	84	492150	49.472	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.940%
36) 1,2-Dichloropropane-d6	5.986	67	166984	52.860	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	105.720%
41) Toluene-d8	7.243	98	410528	45.823	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.640%
43) trans-1,3-Dichloroprop...	7.551	79	65880	42.898	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.800%
47) 2-Hexanone-d5	8.024	63	104676	87.642	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	87.640%
56) 1,1,2,2-Tetrachloroeth...	10.149	84	224042	51.502	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	103.000%
66) 1,2-Dichlorobenzene-d4	11.558	152	149156	51.743	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	103.480%
Target Compounds						
16) Methylene chloride	2.449	84	12132	4.844	ug/L	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040924\
Data File : VV035073.D
Acq On : 09 Apr 2024 11:27
Operator : SY/MD
Sample : P2017-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BH5E9

Quant Time: Apr 10 04:55:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 10 04:53:36 2024
Response via : Initial Calibration

