

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW040924\
 Data File : VW035092.D
 Acq On : 09 Apr 2024 19:03
 Operator : SY/MD
 Sample : P1968-01 10X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E07R9

Quant Time: Apr 10 04:59:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 10 04:53:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.532	114	327482	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	315911	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	127398	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	114443	36.186	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	72.380%
7) Chloroethane-d5	1.523	69	115676	42.110	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	84.220%
11) 1,1-Dichloroethene-d2	2.056	65	53633	38.686	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.380%
21) 2-Butanone-d5	3.786	46	200703	107.001	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.000%
24) Chloroform-d	4.249	84	250088	47.031	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.060%
26) 1,2-Dichloroethane-d4	4.940	65	167677	49.667	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.340%
32) Benzene-d6	4.957	84	474132	44.681	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.360%
36) 1,2-Dichloropropane-d6	5.989	67	165664	49.163	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	98.320%
41) Toluene-d8	7.243	98	394538	41.284	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	82.560%
43) trans-1,3-Dichloroprop...	7.555	79	68876	42.045	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.080%
47) 2-Hexanone-d5	8.024	63	123898	97.249	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	97.250%
56) 1,1,2,2-Tetrachloroeth...	10.153	84	228097	49.155	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	98.320%
66) 1,2-Dichlorobenzene-d4	11.561	152	141818	49.580	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.160%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV040924\
Data File : VV035092.D
Acq On : 09 Apr 2024 19:03
Operator : SY/MD
Sample : P1968-01 10X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E07R9

Quant Time: Apr 10 04:59:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 10 04:53:36 2024
Response via : Initial Calibration

