

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041020\
 Data File : VV015449.D
 Acq On : 10 Apr 2020 12:48
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD00542

Quant Time: Apr 11 04:45:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 10 12:58:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	122590	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	117135	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	60272	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	40377	6.35	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	127.00%
7) Chloroethane-d5	1.58	69	33745	5.76	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	115.20%
11) 1,1-Dichloroethene-d2	2.12	63	76538	5.71	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	114.20%
20) 2-Butanone-d5	3.93	46	87098	48.77	ug/L	-0.03
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.54%
24) Chloroform-d	4.37	84	70668	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.40%
26) 1,2-Dichloroethane-d4	5.06	65	37415	4.93	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.60%
32) Benzene-d6	5.07	84	137838	5.30	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	106.00%
36) 1,2-Dichloropropane-d6	6.09	67	41213	5.09	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.80%
41) Toluene-d8	7.33	98	133053	5.36	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	107.20%
43) trans-1,3-Dichloropropene-	7.64	79	18199	5.17	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	103.40%
46) 2-Hexanone-d5	8.11	63	58295	34.34	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	68.68%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	29854	4.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.00%
64) 1,2-Dichlorobenzene-d4	11.65	152	46112	4.77	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.40%

Target Compounds

					Ovalue
2) Dichlorodifluoromethane	1.14	85	59530	4.906 ug/L	99
3) Chloromethane	1.25	50	54600	4.725 ug/L	100
5) Vinyl chloride	1.32	62	53778	4.833 ug/L	100
6) Bromomethane	1.53	94	31495	5.122 ug/L	99
8) Chloroethane	1.60	64	32665	5.062 ug/L	99
9) Trichlorofluoromethane	1.77	101	73030	5.130 ug/L	99
10) 1,1,2-Trichloro-1,2,2-trif	2.13	101	42188	5.247 ug/L	98
12) 1,1-Dichloroethene	2.13	96	38325	4.980 ug/L	93
13) Acetone	2.21	43	78381	62.118 ug/L #	54
14) Carbon disulfide	2.31	76	122541	4.568 ug/L	100
15) Methyl Acetate	2.46	43	17463	4.678 ug/L	97
16) Methylene chloride	2.52	84	46030	5.106 ug/L	97
17) Methyl tert-butyl Ether	2.79	73	102656	4.991 ug/L	99
18) trans-1,2-Dichloroethene	2.78	96	42370	4.886 ug/L	99
19) 1,1-Dichloroethane	3.21	63	82655	5.056 ug/L	97
21) 2-Butanone	4.01	43	111510	54.686 ug/L #	71
22) cis-1,2-Dichloroethene	3.94	96	45653	5.071 ug/L	92

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041020\
 Data File : VV015449.D
 Acq On : 10 Apr 2020 12:48
 Operator : SY/MD
 Sample : VSTDCCC005
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD00542

Quant Time: Apr 11 04:45:15 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Apr 10 12:58:25 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
23) Bromochloromethane	4.27	128	18837	5.028	ug/L	97
25) Chloroform	4.40	83	85201	5.093	ug/L	99
27) 1,2-Dichloroethane	5.15	62	53292	4.927	ug/L	99
29) 1,1,1-Trichloroethane	4.63	97	76059	5.005	ug/L	99
30) Cyclohexane	4.70	56	78038	4.837	ug/L	98
31) Carbon tetrachloride	4.85	117	65626	4.961	ug/L	99
33) Benzene	5.12	78	174860	4.988	ug/L	100
34) Trichloroethene	5.94	95	46978	4.978	ug/L	98
35) Methylcyclohexane	6.15	83	79336	4.990	ug/L	99
37) 1,2-Dichloropropane	6.20	63	43899	5.049	ug/L	98
38) Bromodichloromethane	6.53	83	57182	4.931	ug/L	95
39) cis-1,3-Dichloropropene	7.05	75	67263	5.041	ug/L	99
40) 4-Methyl-2-pentanone	7.25	43	263078	46.076	ug/L	99
42) Toluene	7.41	91	190731	5.011	ug/L	98
44) trans-1,3-Dichloropropene	7.67	75	56492	5.007	ug/L	96
45) 1,1,2-Trichloroethane	7.86	97	28842	4.934	ug/L	98
47) Tetrachloroethene	7.99	164	34814	4.980	ug/L	98
48) 2-Hexanone	8.16	43	184862	46.339	ug/L	97
49) Dibromochloromethane	8.26	129	35051	5.006	ug/L	98
50) 1,2-Dibromoethane	8.37	107	27157	4.848	ug/L #	100
51) Chlorobenzene	8.90	112	121590	5.012	ug/L	99
52) Ethylbenzene	9.03	91	224197	5.117	ug/L	100
53) m,p-xylene	9.16	106	83185	5.027	ug/L	99
54) o-xylene	9.56	106	80079	5.000	ug/L	98
55) Styrene	9.58	104	134548	4.998	ug/L	99
56) Isopropylbenzene	9.95	105	221589	5.046	ug/L	100
58) 1,1,1,2,2-Tetrachloroethane	10.26	83	34524	4.853	ug/L	99
59) 1,2,3-Trichloropropane	10.29	75	26217	4.876	ug/L	99
61) Bromoform	9.75	173	17073	4.830	ug/L	98
62) 1,3-Dichlorobenzene	11.20	146	97120	4.998	ug/L	97
63) 1,4-Dichlorobenzene	11.29	146	97067	4.909	ug/L	98
65) 1,2-Dichlorobenzene	11.67	146	87683	4.949	ug/L	99
66) 1,2-Dibromo-3-chloropropan	12.45	75	6328	4.668	ug/L	89
67) 1,3,5-Trichlorobenzene	12.67	180	70361	5.011	ug/L	99
68) 1,2,4-trichlorobenzene	13.28	180	60837	4.810	ug/L	99
69) Naphthalene	13.52	128	92726	3.891	ug/L	99
70) 1,2,3-Trichlorobenzene	13.77	180	51117	4.575	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

