

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041020\
 Data File : VV015453.D
 Acq On : 10 Apr 2020 15:28
 Operator : SY/MD
 Sample : L2216-04
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0Y10

Quant Time: Apr 11 06:52:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 11 04:49:20 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	113528	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	108885	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	52004	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	36826	6.25	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	125.00%
7) Chloroethane-d5	1.58	69	32131	5.92	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	118.40%
11) 1,1-Dichloroethene-d2	2.12	63	51969	4.19	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.80%
20) 2-Butanone-d5	3.93	46	83377	50.41	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	100.82%
24) Chloroform-d	4.38	84	64588	4.95	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.00%
26) 1,2-Dichloroethane-d4	5.06	65	35541	5.05	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.00%
32) Benzene-d6	5.08	84	131210	5.43	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.60%
36) 1,2-Dichloropropane-d6	6.09	67	40382	5.37	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	107.40%
41) Toluene-d8	7.34	98	121770	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
43) trans-1,3-Dichloropropene-	7.64	79	16434	5.02	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	100.40%
46) 2-Hexanone-d5	8.11	63	63248	40.08	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.16%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	27150	4.55	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	91.00%
64) 1,2-Dichlorobenzene-d4	11.65	152	43020	5.15	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	103.00%

Target Compounds

					Ovalue
3) Chloromethane	1.25	50	1173	0.110 ug/L	95
9) Trichlorofluoromethane	1.77	101	9088	0.689 ug/L	96
13) Acetone	2.21	43	5621	4.810 ug/L #	53
25) Chloroform	4.41	83	3308	0.214 ug/L	84
34) Trichloroethene	5.94	95	74599	8.505 ug/L	99
47) Tetrachloroethene	8.00	164	8039	1.237 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041020\
Data File : VV015453.D
Acq On : 10 Apr 2020 15:28
Operator : SY/MD
Sample : L2216-04
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0Y10

Quant Time: Apr 11 06:52:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 11 04:49:20 2020
Response via : Initial Calibration

