

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041023\
 Data File : VW030466.D
 Acq On : 10 Apr 2023 14:15
 Operator : SY/MD
 Sample : 02225-12DL 10X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA9DL

Quant Time: Apr 11 06:28:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 11 06:24:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	145796	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	144265	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	69896	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	34945	3.619	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.400%
7) Chloroethane-d5	1.536	69	39595	4.225	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	84.600%
11) 1,1-Dichloroethene-d2	2.066	63	68289	2.980	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	59.600%#
20) 2-Butanone-d5	3.812	46	105374	42.220	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	84.440%
24) Chloroform-d	4.259	84	104902	4.489	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
26) 1,2-Dichloroethane-d4	4.953	65	52225	4.736	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.800%
32) Benzene-d6	4.969	84	184769	4.545	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.000%
36) 1,2-Dichloropropane-d6	5.998	67	59227	4.654	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.000%
41) Toluene-d8	7.252	98	157605	4.222	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.400%
43) trans-1,3-Dichloroprop...	7.561	79	19842	4.141	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.800%
46) 2-Hexanone-d5	8.034	63	81591	42.373	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.740%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	43465	4.333	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.600%
66) 1,2-Dichlorobenzene-d4	11.570	152	63513	5.084	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	101.600%
Target Compounds						
					Qvalue	
9) Trichlorofluoromethane	1.719	101	2364	0.095	ug/L	96
12) 1,1-Dichloroethene	2.076	96	8025	0.659	ug/L #	1
16) Methylene chloride	2.455	84	7698	0.534	ug/L	96
19) 1,1-Dichloroethane	3.121	63	7837	0.315	ug/L #	96
29) 1,1,1-Trichloroethane	4.523	97	155100	6.689	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041023\
Data File : VV030466.D
Acq On : 10 Apr 2023 14:15
Operator : SY/MD
Sample : 02225-12DL 10X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA9DL

Quant Time: Apr 11 06:28:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 11 06:24:59 2023
Response via : Initial Calibration

