

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041023\
 Data File : VW030468.D
 Acq On : 10 Apr 2023 15:05
 Operator : SY/MD
 Sample : 02225-14DL 5X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AB2DL

Quant Time: Apr 11 06:29:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Apr 11 06:24:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	110879	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	110108	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	53605	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	28485	3.879	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	77.600%	
7) Chloroethane-d5	1.536	69	31797	4.462	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	89.200%	
11) 1,1-Dichloroethene-d2	2.063	63	54296	3.116	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.400%	
20) 2-Butanone-d5	3.825	46	85033	44.799	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	89.600%	
24) Chloroform-d	4.259	84	86428	4.863	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	97.200%	
26) 1,2-Dichloroethane-d4	4.953	65	42595	5.079	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	101.600%	
32) Benzene-d6	4.969	84	148286	4.779	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	95.600%	
36) 1,2-Dichloropropane-d6	5.998	67	48495	4.992	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	99.800%	
41) Toluene-d8	7.252	98	125902	4.419	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	88.400%	
43) trans-1,3-Dichloroprop...	7.561	79	15457	4.226	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	84.600%	
46) 2-Hexanone-d5	8.037	63	66452	45.216	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	90.440%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	36030	4.706	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	94.200%	
66) 1,2-Dichlorobenzene-d4	11.570	152	50794	5.302	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	106.000%	
Target Compounds						
					Qvalue	
8) Chloroethane	1.552	64	1227	0.165	ug/L	84
16) Methylene chloride	2.455	84	10516	0.960	ug/L	98
19) 1,1-Dichloroethane	3.117	63	133415	7.051	ug/L	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041023\
Data File : VV030468.D
Acq On : 10 Apr 2023 15:05
Operator : SY/MD
Sample : 02225-14DL 5X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AB2DL

Quant Time: Apr 11 06:29:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Apr 11 06:24:59 2023
Response via : Initial Calibration

