

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041024\
 Data File : VW035109.D
 Acq On : 10 Apr 2024 15:16
 Operator : SY/MD
 Sample : P2006-15DL 500X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :

Quant Time: Apr 11 06:58:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Apr 10 04:53:36 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	324050	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.783	117	312727	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	139381	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	164915	52.697	ug/L	0.00
Spiked Amount 50.000	Range 60 - 135		Recovery = 105.400%			
7) Chloroethane-d5	1.532	69	152595	56.138	ug/L	0.00
Spiked Amount 50.000	Range 70 - 130		Recovery = 112.280%			
11) 1,1-Dichloroethene-d2	2.057	65	68814	50.162	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 100.320%			
21) 2-Butanone-d5	3.793	46	204509	110.185	ug/L	0.00
Spiked Amount 100.000	Range 40 - 130		Recovery = 110.180%			
24) Chloroform-d	4.246	84	271716	51.640	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 103.280%			
26) 1,2-Dichloroethane-d4	4.941	65	181423	54.308	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 108.620%			
32) Benzene-d6	4.957	84	549708	52.330	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery = 104.660%			
36) 1,2-Dichloropropane-d6	5.986	67	182087	54.587	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery = 109.180%			
41) Toluene-d8	7.243	98	466892	49.353	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 98.700%			
43) trans-1,3-Dichloroprop...	7.551	79	74655	46.036	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery = 92.080%			
47) 2-Hexanone-d5	8.027	63	117309	93.015	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery = 93.010%			
56) 1,1,2,2-Tetrachloroeth...	10.153	84	236618	51.511	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery = 103.020%			
66) 1,2-Dichlorobenzene-d4	11.558	152	170827	54.588	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery = 109.180%			
Target Compounds						Qvalue
10) 1,1,2-Trichloro-1,2,2-...	2.057	101	1975	0.777	ug/L #	22
33) Benzene	5.011	78	136808	13.416	ug/L	100
48) 2-Hexanone	8.088	43	14853	4.854	ug/L #	77
51) Chlorobenzene	8.812	112	402616	58.186	ug/L	99
64) 1,3-Dichlorobenzene	11.124	146	20527	4.371	ug/L	95
65) 1,4-Dichlorobenzene	11.207	146	161889	33.902	ug/L	99
67) 1,2-Dichlorobenzene	11.577	146	174090	37.216	ug/L	98
69) 1,3,5-Trichlorobenzene	13.201	180	27428	8.105	ug/L	98
70) 1,2,4-trichlorobenzene	13.201	180	27428	8.823	ug/L	99
72) 1,2,3-Trichlorobenzene	13.683	180	8630	2.816	ug/L	95


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041024\
Data File : VV035109.D
Acq On : 10 Apr 2024 15:16
Operator : SY/MD
Sample : P2006-15DL 500X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :

Quant Time: Apr 11 06:58:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM032624WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Apr 10 04:53:36 2024
Response via : Initial Calibration









#10
1,1,2-Trichloro-1,2,2-trifluoroethane
Concen: 0.777 ug/L
RT: 2.057 min Scan# 31
Delta R.T. -0.010 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16




Instrument :
MSVOA_V
ClientSampleId :


Tgt Ion:101 Resp: 1975
Ion Ratio Lower Upper
101 100
85 0.0 35.3 52.9#
151 0.0 55.0 82.4#



#33
Benzene
Concen: 13.416 ug/L
RT: 5.011 min Scan# 1235
Delta R.T. 0.007 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16


Tgt Ion: 78 Resp: 136808








#48
2-Hexanone
Concen: 4.854 ug/L
RT: 8.088 min Scan# 2187
Delta R.T. 0.016 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16

Instrument :
MSVOA_V
ClientSampleId :



Tgt Ion: 43 Resp: 14853




Ion	Ratio	Lower	Upper
43	100		
58	44.5	45.4	68.2#
57	0.0	16.2	24.4#
100	5.9	10.1	15.1#




#51
Chlorobenzene
Concen: 58.186 ug/L
RT: 8.812 min Scan# 2417
Delta R.T. 0.000 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16

Tgt Ion: 112 Resp: 402616


Ion	Ratio	Lower	Upper
112	100		
114	32.6	22.0	41.0
77	60.1	47.4	71.2








#64
1,3-Dichlorobenzene
Concen: 4.371 ug/L
RT: 11.124 min Scan# 3133
Delta R.T. 0.010 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16

Instrument :
MSVOA_V
ClientSampleId :







Tgt Ion:146 Resp: 20527
Ion Ratio Lower Upper
146 100
111 35.5 28.7 53.3
148 61.3 44.0 81.8



#65
1,4-Dichlorobenzene
Concen: 33.902 ug/L
RT: 11.207 min Scan# 3162
Delta R.T. 0.000 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16

Tgt Ion:146 Resp: 161889
Ion Ratio Lower Upper
146 100
111 39.8 27.4 50.8
148 63.4 44.2 82.2









#67
1,2-Dichlorobenzene
Concen: 37.216 ug/L
RT: 11.577 min Scan# 31
Delta R.T. 0.000 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16




Instrument :
MSVOA_V
ClientSampleId :


Tgt Ion:146 Resp: 174090
Ion Ratio Lower Upper
146 100
111 41.6 34.6 52.0
148 63.8 51.5 77.3



#69
1,3,5-Trichlorobenzene
Concen: 8.105 ug/L
RT: 13.201 min Scan# 3782
Delta R.T. 0.621 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16


Tgt Ion:180 Resp: 27428
Ion Ratio Lower Upper
180 100
182 95.6 74.7 112.1
184 31.3 24.0 36.0
145 30.7 23.6 35.4








#70
1,2,4-trichlorobenzene
Concen: 8.823 ug/L
RT: 13.201 min Scan# 31
Delta R.T. 0.007 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16

Instrument :
MSVOA_V
ClientSampleId :



Tgt Ion:180 Resp: 27428
Ion Ratio Lower Upper
180 100
182 95.6 75.6 113.4
145 30.7 24.1 36.1



#72
1,2,3-Trichlorobenzene
Concen: 2.816 ug/L
RT: 13.683 min Scan# 3932
Delta R.T. 0.007 min
Lab File: VV035109.D
Acq: 10 Apr 2024 15:16

Tgt Ion:180 Resp: 8630
Ion Ratio Lower Upper
180 100
182 89.3 76.2 114.2
145 31.3 25.8 38.6

