

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041123\
 Data File : VW030500.D
 Acq On : 11 Apr 2023 17:25
 Operator : SY/MD
 Sample : 02275-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHDY9

Quant Time: Apr 12 05:04:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Apr 12 05:00:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	107565	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.793	117	106533	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	51988	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.282	65	22659	3.181	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	63.600%
7) Chloroethane-d5	1.536	69	27067	3.915	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	78.400%
11) 1,1-Dichloroethene-d2	2.063	63	44549	2.635	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	52.800%#
20) 2-Butanone-d5	3.828	46	79863	43.372	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	86.740%
24) Chloroform-d	4.259	84	74407	4.315	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.400%
26) 1,2-Dichloroethane-d4	4.954	65	38173	4.692	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.800%
32) Benzene-d6	4.970	84	125281	4.173	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.400%
36) 1,2-Dichloropropane-d6	5.998	67	40790	4.340	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.800%
41) Toluene-d8	7.252	98	104134	3.777	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.600%
43) trans-1,3-Dichloroprop...	7.564	79	14068	3.976	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.600%
46) 2-Hexanone-d5	8.037	63	60557	42.588	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.180%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	32522	4.390	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.800%
66) 1,2-Dichlorobenzene-d4	11.571	152	43780	4.712	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	94.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041123\
Data File : VV030500.D
Acq On : 11 Apr 2023 17:25
Operator : SY/MD
Sample : 02275-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHDY9

Quant Time: Apr 12 05:04:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Apr 12 05:00:18 2023
Response via : Initial Calibration

