

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041124\
 Data File : VW035145.D
 Acq On : 12 Apr 2024 00:30
 Operator : SY/MD
 Sample : P2057-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COBL2

Quant Time: Apr 12 02:07:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 12 02:02:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	169440	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	173565	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.184	152	75716	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	54837	3.927	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.600%
7) Chloroethane-d5	1.532	69	55852	4.851	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.000%
11) 1,1-Dichloroethene-d2	2.059	65	26414	4.165	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	83.200%
20) 2-Butanone-d5	3.812	46	142235	65.099	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	130.200%#
24) Chloroform-d	4.249	84	117875	5.217	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.400%
26) 1,2-Dichloroethane-d4	4.947	65	60239	5.904	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	118.000%
32) Benzene-d6	4.960	84	225888	4.975	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.600%
36) 1,2-Dichloropropane-d6	5.989	67	72948	5.474	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	109.400%
41) Toluene-d8	7.246	98	179934	4.363	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.200%
43) trans-1,3-Dichloroprop...	7.561	79	20376	4.304	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	86.000%
46) 2-Hexanone-d5	8.027	63	101189	62.283	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	124.560%
56) 1,1,2,2-Tetrachloroeth...	10.152	84	56431	6.069	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	121.400%#
66) 1,2-Dichlorobenzene-d4	11.561	152	67561	5.665	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	113.200%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041124\
Data File : VV035145.D
Acq On : 12 Apr 2024 00:30
Operator : SY/MD
Sample : P2057-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BL2

Quant Time: Apr 12 02:07:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 12 02:02:37 2024
Response via : Initial Calibration

