

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041124\
 Data File : VW035146.D
 Acq On : 12 Apr 2024 00:54
 Operator : SY/MD
 Sample : P2057-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COBL4

Quant Time: Apr 12 02:07:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040224WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 12 02:02:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	171551	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.786	117	172079	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.185	152	76619	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	59510	4.209	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	84.200%	
7) Chloroethane-d5	1.532	69	58979	5.060	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	101.200%	
11) 1,1-Dichloroethene-d2	2.060	65	27590	4.297	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	86.000%	
20) 2-Butanone-d5	3.812	46	153306	69.303	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	138.600%#	
24) Chloroform-d	4.249	84	125654	5.493	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	109.800%	
26) 1,2-Dichloroethane-d4	4.944	65	62364	6.037	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	120.800%	
32) Benzene-d6	4.960	84	242339	5.384	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	107.600%	
36) 1,2-Dichloropropane-d6	5.989	67	80587	6.100	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	122.000%	
41) Toluene-d8	7.243	98	196531	4.806	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	96.200%	
43) trans-1,3-Dichloroprop...	7.561	79	19661	4.189	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	83.800%	
46) 2-Hexanone-d5	8.027	63	129089	80.142	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	160.280%#	
56) 1,1,2,2-Tetrachloroeth...	10.153	84	59398	6.444	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	128.800%#	
66) 1,2-Dichlorobenzene-d4	11.561	152	72141	5.978	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	119.600%	
Target Compounds						Qvalue
25) Chloroform	4.278	83	8846	0.350	ug/L	94
34) Trichloroethene	5.841	95	26228	1.911	ug/L	95
47) Tetrachloroethene	7.918	164	21959023	2129.314	ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041124\
Data File : VV035146.D
Acq On : 12 Apr 2024 00:54
Operator : SY/MD
Sample : P2057-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
COBL4

Quant Time: Apr 12 02:07:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR040224WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 12 02:02:37 2024
Response via : Initial Calibration

