

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041223\
 Data File : VW030519.D
 Acq On : 12 Apr 2023 15:45
 Operator : SY/MD
 Sample : 02254-06DL 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AA8DL

Quant Time: Apr 13 02:49:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 13 02:44:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	109254	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	108675	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	51464	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	28538	3.944	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	78.800%
7) Chloroethane-d5	1.535	69	26640	3.794	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	75.800%
11) 1,1-Dichloroethene-d2	2.063	63	72147	4.202	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	84.000%
20) 2-Butanone-d5	3.841	46	68620	36.690	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	73.380%
24) Chloroform-d	4.259	84	72620	4.147	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	83.000%
26) 1,2-Dichloroethane-d4	4.956	65	38185	4.621	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.400%
32) Benzene-d6	4.969	84	128909	4.209	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.200%
36) 1,2-Dichloropropane-d6	6.001	67	41473	4.326	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	86.600%
41) Toluene-d8	7.252	98	112005	3.983	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	79.600%
43) trans-1,3-Dichloroprop...	7.564	79	14590	4.042	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.800%
46) 2-Hexanone-d5	8.040	63	59023	40.691	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	81.380%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	31503	4.169	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	83.400%
66) 1,2-Dichlorobenzene-d4	11.570	152	44662	4.855	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.200%
Target Compounds						
12) 1,1-Dichloroethene	2.072	96	34292	3.757	ug/L	97
19) 1,1-Dichloroethane	3.117	63	98860	5.303	ug/L	98
29) 1,1,1-Trichloroethane	4.522	97	39768	2.277	ug/L	99
34) Trichloroethene	5.850	95	1941	0.186	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041223\
Data File : VV030519.D
Acq On : 12 Apr 2023 15:45
Operator : SY/MD
Sample : 02254-06DL 20X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AA8DL

Quant Time: Apr 13 02:49:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 13 02:44:31 2023
Response via : Initial Calibration

