

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041223\
 Data File : VW030522.D
 Acq On : 12 Apr 2023 17:00
 Operator : SY/MD
 Sample : 02250-03DL 20X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHDx4DL

Quant Time: Apr 13 02:50:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Apr 13 02:44:31 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	99274	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.796	117	100158	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	48663	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	48686	7.405	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	148.000%#
7) Chloroethane-d5	1.535	69	37429	5.866	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	117.400%
11) 1,1-Dichloroethene-d2	2.063	63	71289	4.569	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	91.400%
20) 2-Butanone-d5	3.812	46	84306	49.608	ug/L	-0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.220%
24) Chloroform-d	4.262	84	80737	5.073	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.400%
26) 1,2-Dichloroethane-d4	4.953	65	41762	5.562	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	111.200%
32) Benzene-d6	4.973	84	145397	5.152	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.000%
36) 1,2-Dichloropropane-d6	5.998	67	45749	5.177	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	103.600%
41) Toluene-d8	7.252	98	127130	4.905	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	98.000%
43) trans-1,3-Dichloroprop...	7.564	79	16858	5.067	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.400%
46) 2-Hexanone-d5	8.034	63	65751	49.184	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	98.360%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	33036	4.744	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.800%
66) 1,2-Dichlorobenzene-d4	11.570	152	47892	5.506	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	110.200%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.076	96	4903	0.591	ug/L #	1
14) Carbon disulfide	2.246	76	2943	0.108	ug/L	96
19) 1,1-Dichloroethane	3.121	63	3620	0.214	ug/L #	90
22) cis-1,2-Dichloroethene	3.828	96	4599	0.524	ug/L	99
34) Trichloroethene	5.844	95	53518	5.578	ug/L	98
47) Tetrachloroethene	7.915	164	13537	1.599	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041223\
Data File : VV030522.D
Acq On : 12 Apr 2023 17:00
Operator : SY/MD
Sample : 02250-03DL 20X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHDx4DL

Quant Time: Apr 13 02:50:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Apr 13 02:44:31 2023
Response via : Initial Calibration

