

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041319\
 Data File : VV010324.D
 Acq On : 13 Apr 2019 17:52
 Operator : SY/MD
 Sample : VV0413WBL01
 Misc : 25mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK35

Quant Time: Apr 14 01:20:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR041319WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Sat Apr 13 14:26:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	137426	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	133531	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	61822	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	34222	4.68	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.60%
7) Chloroethane-d5	1.58	69	39634	4.94	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.80%
11) 1,1-Dichloroethene-d2	2.13	63	72060	3.53	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	70.60%
20) 2-Butanone-d5	3.96	46	99343	53.24	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	106.48%
24) Chloroform-d	4.40	84	84354	4.99	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	99.80%
26) 1,2-Dichloroethane-d4	5.09	65	44845	5.26	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.20%
32) Benzene-d6	5.10	84	161632	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.20%
36) 1,2-Dichloropropane-d6	6.12	67	49750	5.10	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.00%
41) Toluene-d8	7.36	98	141911	4.76	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.20%
43) trans-1,3-Dichloropropene-	7.67	79	15717	5.05	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	101.00%
46) 2-Hexanone-d5	8.14	63	74583	52.95	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	105.90%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	39488	5.28	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	105.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	59662	5.35	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	107.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041319\
Data File : VV010324.D
Acq On : 13 Apr 2019 17:52
Operator : SY/MD
Sample : VV0413WBL01
Misc : 25mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK35

Quant Time: Apr 14 01:20:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR041319WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Sat Apr 13 14:26:57 2019
Response via : Initial Calibration

