

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041323\
 Data File : VW030538.D
 Acq On : 13 Apr 2023 12:24
 Operator : SY/MD
 Sample : 02250-05DL 50X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHD6DL

Quant Time: Apr 14 05:19:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 14 05:16:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	127005	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.796	117	126909	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	58924	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	30263	3.598	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	72.000%
7) Chloroethane-d5	1.535	69	31368	3.843	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	76.800%
11) 1,1-Dichloroethene-d2	2.066	63	86831	4.350	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	87.000%
20) 2-Butanone-d5	3.831	46	83255	38.293	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	76.580%
24) Chloroform-d	4.262	84	82958	4.075	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	81.400%
26) 1,2-Dichloroethane-d4	4.956	65	43084	4.485	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.800%
32) Benzene-d6	4.973	84	143568	4.015	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.200%
36) 1,2-Dichloropropane-d6	5.998	67	47305	4.225	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.600%
41) Toluene-d8	7.252	98	123950	3.774	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	7.564	79	15821	3.753	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	75.000%
46) 2-Hexanone-d5	8.037	63	66676	39.362	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	78.720%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	35024	3.969	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	79.400%
66) 1,2-Dichlorobenzene-d4	11.570	152	48551	4.610	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	92.200%
Target Compounds						
10) 1,1,2-Trichloro-1,2,2-...	2.072	101	4614	0.386	ug/L	99
12) 1,1-Dichloroethene	2.076	96	46815	4.412	ug/L	91
19) 1,1-Dichloroethane	3.124	63	29954	1.382	ug/L	98
22) cis-1,2-Dichloroethene	3.828	96	3019	0.269	ug/L	92
29) 1,1,1-Trichloroethane	4.526	97	7749	0.380	ug/L	89
34) Trichloroethene	5.844	95	70586	5.806	ug/L	98
47) Tetrachloroethene	7.918	164	12318	1.148	ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041323\
Data File : VV030538.D
Acq On : 13 Apr 2023 12:24
Operator : SY/MD
Sample : 02250-05DL 50X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHDx6DL

Quant Time: Apr 14 05:19:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 14 05:16:37 2023
Response via : Initial Calibration

