

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041323\
 Data File : VW030543.D
 Acq On : 13 Apr 2023 14:19
 Operator : SY/MD
 Sample : 02254-15
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AD9

Quant Time: Apr 14 05:20:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Apr 14 05:16:37 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	117197	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.792	117	115031	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	53903	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.281	65	46557	5.998	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	120.000%
7) Chloroethane-d5	1.535	69	35472	4.709	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.200%
11) 1,1-Dichloroethene-d2	2.063	63	68914	3.742	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.800%
20) 2-Butanone-d5	3.828	46	82670	41.206	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	82.420%
24) Chloroform-d	4.262	84	81534	4.340	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.800%
26) 1,2-Dichloroethane-d4	4.956	65	42915	4.842	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.800%
32) Benzene-d6	4.973	84	144404	4.455	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.000%
36) 1,2-Dichloropropane-d6	6.001	67	44632	4.398	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.000%
41) Toluene-d8	7.252	98	121985	4.098	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.000%
43) trans-1,3-Dichloroprop...	7.564	79	16673	4.364	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.200%
46) 2-Hexanone-d5	8.037	63	64304	41.882	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.760%
56) 1,1,2,2-Tetrachloroeth...	10.162	84	34173	4.273	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.400%
66) 1,2-Dichlorobenzene-d4	11.570	152	46887	4.867	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.400%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041323\
Data File : VV030543.D
Acq On : 13 Apr 2023 14:19
Operator : SY/MD
Sample : 02254-15
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AD9

Quant Time: Apr 14 05:20:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Apr 14 05:16:37 2023
Response via : Initial Calibration

