

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041420\
 Data File : VV015481.D
 Acq On : 14 Apr 2020 23:27
 Operator : SY/MD
 Sample : L2274-01
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YAQ99

Quant Time: Apr 15 05:52:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 15 05:31:01 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	109505	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	105125	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	51102	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	32489	5.72	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	114.40%
7) Chloroethane-d5	1.58	69	28183	5.39	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	107.80%
11) 1,1-Dichloroethene-d2	2.12	63	45685	3.82	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	76.40%
20) 2-Butanone-d5	3.93	46	78092	48.95	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.90%
24) Chloroform-d	4.38	84	57499	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.40%
26) 1,2-Dichloroethane-d4	5.06	65	33672	4.96	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.20%
32) Benzene-d6	5.08	84	121842	5.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.40%
36) 1,2-Dichloropropane-d6	6.10	67	36899	5.08	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	101.60%
41) Toluene-d8	7.34	98	112123	5.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.80%
43) trans-1,3-Dichloropropene-	7.65	79	14697	4.65	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	93.00%
46) 2-Hexanone-d5	8.11	63	63795	41.87	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.74%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	25878	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.80%
64) 1,2-Dichlorobenzene-d4	11.65	152	40587	4.95	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.00%

Target Compounds					Ovalue
13) Acetone	2.22	43	6458	5.730 ug/L	94
25) Chloroform	4.41	83	18840	1.261 ug/L	94
47) Tetrachloroethene	8.00	164	20723	3.303 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041420\
Data File : VV015481.D
Acq On : 14 Apr 2020 23:27
Operator : SY/MD
Sample : L2274-01
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
Yaq99

Quant Time: Apr 15 05:52:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 15 05:31:01 2020
Response via : Initial Calibration

