

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041420\
 Data File : VV015491.D
 Acq On : 15 Apr 2020 03:21
 Operator : SY/MD
 Sample : L2274-11
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YQA9

Quant Time: Apr 15 06:53:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Apr 15 05:31:01 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	106213	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	102371	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	50646	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	31752	5.76	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.20%
7) Chloroethane-d5	1.58	69	28128	5.54	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	110.80%
11) 1,1-Dichloroethene-d2	2.12	63	44785	3.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.20%
20) 2-Butanone-d5	3.93	46	77022	49.78	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.56%
24) Chloroform-d	4.38	84	59143	4.85	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	97.00%
26) 1,2-Dichloroethane-d4	5.06	65	33736	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
32) Benzene-d6	5.08	84	119690	5.27	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	105.40%
36) 1,2-Dichloropropane-d6	6.10	67	36907	5.22	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	104.40%
41) Toluene-d8	7.34	98	111155	5.13	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.60%
43) trans-1,3-Dichloropropene-	7.65	79	14703	4.78	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	95.60%
46) 2-Hexanone-d5	8.11	63	61543	41.48	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	82.96%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	25829	4.60	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	92.00%
64) 1,2-Dichlorobenzene-d4	11.65	152	40608	4.99	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	99.80%

Target Compounds

					Ovalue
2) Dichlorodifluoromethane	1.14	85	1315	0.125 ug/L	94
13) Acetone	2.21	43	9040	8.269 ug/L #	53
25) Chloroform	4.41	83	60053	4.143 ug/L	98
47) Tetrachloroethene	8.00	164	1845	0.302 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041420\
Data File : VV015491.D
Acq On : 15 Apr 2020 03:21
Operator : SY/MD
Sample : L2274-11
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YQA9

Quant Time: Apr 15 06:53:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Apr 15 05:31:01 2020
Response via : Initial Calibration

