


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041421\
Data File : VV021014.D
Acq On : 14 Apr 2021 18:34
Operator : SY/MD
Sample : M2031-06
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG536

Manual Integrations
APPROVED

MMDadoda
4/16/2021 1:03:27 PM

Quant Time: Apr 15 03:06:54 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041421WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Apr 15 01:36:39 2021
Response via : Initial Calibration



Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041421\
 Data File : VV021014.D
 Acq On : 14 Apr 2021 18:34
 Operator : SY/MD
 Sample : M2031-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG536

Manual Integrations
 APPROVED

MMDadoda
 4/16/2021 1:03:27 PM

Quant Time: Apr 15 03:06:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041421WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 15 01:36:39 2021
 Response via : Initial Calibration



(13) Acetone (T)

2.182min (-0.009) 1.12 ug/L

response 1491

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	30.90	47.02
0.00	0.00	0.00
0.00	0.00	0.00

Quantitation Report (Qedit)




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041421\
 Data File : VV021014.D
 Acq On : 14 Apr 2021 18:34
 Operator : SY/MD
 Sample : M2031-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG536

Manual Integrations
 APPROVED

MMDadoda
 4/16/2021 1:03:27 PM

Quant Time: Apr 15 03:06:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041421WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 15 01:36:39 2021
 Response via : Initial Calibration



(13) Acetone (T)

2.195min (+ 0.004) 3.09 ug/L m

response 4095

Ion	Exp%	Act%
43.00	100.00	100.00
58.00	30.90	17.12
0.00	0.00	0.00
0.00	0.00	0.00

> MB
 4/16/21

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041421\
 Data File : VW021014.D
 Acq On : 14 Apr 2021 18:34
 Operator : SY/MD
 Sample : M2031-06
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG536

Manual Integrations
 APPROVED

MMDadoda
 4/16/2021 1:03:27 PM

Quant Time: Apr 15 03:06:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM041421WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Apr 15 01:36:39 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.622	114	442752	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.857	117	423280	50.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.255	152	197493	50.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.310	65	123255	51.17	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	102.34%	
7) Chloroethane-d5	1.571	69	101047	53.37	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	106.74%	
11) 1,1-Dichloroethene-d2	2.111	63	198460	38.41	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	76.82%	
21) 2-Butanone-d5	3.892	46	183930	116.11	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	116.11%	
24) Chloroform-d	4.352	84	315086	52.70	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	105.40%	
26) 1,2-Dichloroethane-d4	5.037	65	224001	55.43	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	110.86%	
32) Benzene-d6	5.056	84	569334	54.48	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	108.96%	
36) 1,2-Dichloropropane-d6	6.072	67	162637	56.06	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	112.12%	
41) Toluene-d8	7.320	98	553019	53.40	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	106.80%	
43) trans-1,3-Dichloroprop...	7.625	79	94423	52.81	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	105.62%	
47) 2-Hexanone-d5	8.092	63	167903	134.73	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	134.73%#	
56) 1,1,2,2-Tetrachloroeth...	10.220	84	263761	56.09	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	112.18%	
66) 1,2-Dichlorobenzene-d4	11.632	152	241095	59.80	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	119.60%	
Target Compounds						Qvalue
13) Acetone	2.195	43	4095m	3.09	ug/L	
25) Chloroform	4.381	83	12401	2.01	ug/L	96

> MD
 4/16/21

(#) = qualifier out of range (m) = manual integration (+) = signals summed