

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW041423\
 Data File : VW030569.D
 Acq On : 14 Apr 2023 13:37
 Operator : SY/MD
 Sample : 02275-13DL 100X
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BHDZ8DL

Quant Time: Apr 15 01:59:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Apr 15 01:54:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.545	114	122614	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.796	117	117137	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.194	152	55104	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	44407	5.468	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	109.400%	
7) Chloroethane-d5	1.536	69	34143	4.332	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	86.600%	
11) 1,1-Dichloroethene-d2	2.063	63	68397	3.549	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	71.000%	
20) 2-Butanone-d5	3.834	46	80106	38.164	ug/L	0.02
Spiked Amount	50.000	Range 40 - 130	Recovery	=	76.320%	
24) Chloroform-d	4.262	84	79575	4.049	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.000%	
26) 1,2-Dichloroethane-d4	4.957	65	43294	4.669	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.400%	
32) Benzene-d6	4.973	84	139138	4.215	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	84.400%	
36) 1,2-Dichloropropane-d6	5.998	67	44143	4.272	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	85.400%	
41) Toluene-d8	7.252	98	118501	3.909	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	78.200%	
43) trans-1,3-Dichloroprop...	7.564	79	14914	3.833	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	76.600%	
46) 2-Hexanone-d5	8.037	63	64265	41.104	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	82.200%	
56) 1,1,2,2-Tetrachloroeth...	10.162	84	33333	4.093	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	81.800%	
66) 1,2-Dichlorobenzene-d4	11.570	152	44429	4.511	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	90.200%	
Target Compounds						
5) Vinyl chloride	1.285	62	1749	0.122	ug/L #	29
12) 1,1-Dichloroethene	2.069	96	2540	0.248	ug/L #	1
22) cis-1,2-Dichloroethene	3.828	96	22473	2.075	ug/L #	95
34) Trichloroethene	5.841	95	75215	6.703	ug/L	98
47) Tetrachloroethene	7.915	164	26979	2.724	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV041423\
Data File : VV030569.D
Acq On : 14 Apr 2023 13:37
Operator : SY/MD
Sample : 02275-13DL 100X
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BHDZ8DL

Quant Time: Apr 15 01:59:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR032323WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Apr 15 01:54:33 2023
Response via : Initial Calibration

