

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041520\  
 Data File : VV015504.D  
 Acq On : 15 Apr 2020 13:02  
 Operator : SY/MD  
 Sample : VV0415WBL01  
 Misc : 25.0mL/MSVOA V/WATER  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VBLK37

Quant Time: Apr 16 00:55:16 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR040920WMA.M  
 Quant Title : TRACE VOA SOM01.0  
 QLast Update : Thu Apr 16 00:42:12 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc | Units | Dev(Min) |
|----------------------------|-------|------|----------|------|-------|----------|
| 1) 1,4-Difluorobenzene     | 5.64  | 114  | 112340   | 5.00 | ug/L  | 0.00     |
| 28) Chlorobenzene-d5       | 8.87  | 117  | 104634   | 5.00 | ug/L  | 0.00     |
| 60) 1,4-Dichlorobenzene-d4 | 11.27 | 152  | 50856    | 5.00 | ug/L  | 0.00     |

## System Monitoring Compounds

|                                |        |       |          |          |      |         |
|--------------------------------|--------|-------|----------|----------|------|---------|
| 4) Vinyl Chloride-d3           | 1.31   | 65    | 29430    | 5.05     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 40 - 130 | Recovery | =    | 101.00% |
| 7) Chloroethane-d5             | 1.58   | 69    | 26717    | 4.98     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 130 | Recovery | =    | 99.60%  |
| 11) 1,1-Dichloroethene-d2      | 2.12   | 63    | 41357    | 3.37     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 125 | Recovery | =    | 67.40%  |
| 20) 2-Butanone-d5              | 3.92   | 46    | 78005    | 47.66    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 40 - 130 | Recovery | =    | 95.32%  |
| 24) Chloroform-d               | 4.38   | 84    | 58076    | 4.50     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 90.00%  |
| 26) 1,2-Dichloroethane-d4      | 5.06   | 65    | 32155    | 4.62     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 92.40%  |
| 32) Benzene-d6                 | 5.08   | 84    | 115773   | 4.99     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 125 | Recovery | =    | 99.80%  |
| 36) 1,2-Dichloropropane-d6     | 6.10   | 67    | 36268    | 5.02     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 60 - 140 | Recovery | =    | 100.40% |
| 41) Toluene-d8                 | 7.34   | 98    | 105966   | 4.78     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 70 - 130 | Recovery | =    | 95.60%  |
| 43) trans-1,3-Dichloropropene- | 7.64   | 79    | 14424    | 4.59     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 55 - 130 | Recovery | =    | 91.80%  |
| 46) 2-Hexanone-d5              | 8.11   | 63    | 62084    | 40.94    | ug/L | 0.00    |
| Spiked Amount                  | 50.000 | Range | 45 - 130 | Recovery | =    | 81.88%  |
| 57) 1,1,2,2-Tetrachloroethane- | 10.24  | 84    | 24220    | 4.22     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 65 - 120 | Recovery | =    | 84.40%  |
| 64) 1,2-Dichlorobenzene-d4     | 11.65  | 152   | 38844    | 4.76     | ug/L | 0.00    |
| Spiked Amount                  | 5.000  | Range | 80 - 120 | Recovery | =    | 95.20%  |

| Target Compounds | Ovalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041520\  
Data File : VV015504.D  
Acq On : 15 Apr 2020 13:02  
Operator : SY/MD  
Sample : VV0415WBL01  
Misc : 25.0mL/MSVOA V/WATER  
ALS Vial : 4 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VBLK37

Quant Time: Apr 16 00:55:16 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_V\METHOD\SOMVTR040920WMA.M  
Quant Title : TRACE VOA SOM01.0  
QLast Update : Thu Apr 16 00:42:12 2020  
Response via : Initial Calibration

