

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV041520\
 Data File : VV015507.D
 Acq On : 15 Apr 2020 14:25
 Operator : SY/MD
 Sample : L2273-03DL 10X
 Misc : 25.0mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 YAQ81DL

Quant Time: Apr 16 01:27:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Thu Apr 16 00:42:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	112227	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	105593	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	51111	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	30258	5.19	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	103.80%
7) Chloroethane-d5	1.58	69	27319	5.10	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	102.00%
11) 1,1-Dichloroethene-d2	2.12	63	43750	3.57	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.40%
20) 2-Butanone-d5	3.93	46	79208	48.44	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	96.88%
24) Chloroform-d	4.38	84	59416	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.20%
26) 1,2-Dichloroethane-d4	5.06	65	33354	4.80	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.00%
32) Benzene-d6	5.08	84	118633	5.06	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.20%
36) 1,2-Dichloropropane-d6	6.10	67	36500	5.00	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	100.00%
41) Toluene-d8	7.34	98	108091	4.83	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	96.60%
43) trans-1,3-Dichloropropene-	7.65	79	14991	4.72	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	94.40%
46) 2-Hexanone-d5	8.11	63	63911	41.76	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.52%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	25340	4.38	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	87.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	39587	4.83	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	96.60%

Target Compounds					Ovalue
25) Chloroform	4.41	83	4371	0.285 ug/L	100
47) Tetrachloroethene	8.00	164	70572	11.198 ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV041520\
Data File : VV015507.D
Acq On : 15 Apr 2020 14:25
Operator : SY/MD
Sample : L2273-03DL 10X
Misc : 25.0mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
YAQ81DL

Quant Time: Apr 16 01:27:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR040920WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Thu Apr 16 00:42:12 2020
Response via : Initial Calibration

